

Keysight 42030A Four-Terminal Pair Standard Resistor Set

Operation and
Service Manual

Notices

© Keysight Technologies
1992 - 2019

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

42030-90001

Edition

Edition 3, August 28, 2019

Printed in Malaysia

Published by:

Keysight Technologies International
Japan G.K.,
1-3-3 Higashikawasaki-cho
Chuo-ku
Kobe-shi, Hyogo, Japan

Warranty

THE MATERIAL CONTAINED IN THIS DOCUMENT IS PROVIDED "AS IS," AND IS SUBJECT TO BEING CHANGED, WITHOUT NOTICE, IN FUTURE EDITIONS. FURTHER, TO THE MAXIMUM EXTENT PERMITTED BY APPLICABLE LAW, KEYSIGHT DISCLAIMS ALL WARRANTIES, EITHER EXPRESS OR IMPLIED WITH REGARD TO THIS MANUAL AND ANY INFORMATION CONTAINED HEREIN, INCLUDING BUT NOT LIMITED TO THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE. KEYSIGHT SHALL NOT BE LIABLE FOR ERRORS OR FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH THE FURNISHING, USE, OR PERFORMANCE OF THIS DOCUMENT OR ANY INFORMATION CONTAINED HEREIN. SHOULD KEYSIGHT AND THE USER HAVE A SEPARATE WRITTEN AGREEMENT WITH WARRANTY TERMS COVERING THE MATERIAL IN THIS DOCUMENT THAT CONFLICT WITH THESE TERMS, THE WARRANTY TERMS IN THE SEPA-

RATE AGREEMENT WILL CONTROL.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Declaration of Conformity

Declarations of Conformity for this product and for other Keysight products may be downloaded from the Web. Go to <http://www.keysight.com/go/conformity>. You can then search by product number to find the latest Declaration of Conformity.

U.S. Government Rights

The Software is "commercial computer software," as defined by Federal Acquisition Regulation ("FAR") 2.101. Pursuant to FAR 12.212 and 27.405-3 and Department of Defense FAR Supplement ("DFARS") 227.7202, the U.S. government acquires commercial computer software under the same terms by which the software is customarily provided to the public. Accordingly, Keysight provides the Software to U.S. government customers under its standard commercial license, which is embodied in its End User License Agreement (EULA), a copy of which can be found at <http://www.keysight.com/find/sweula>. The license set forth in the EULA represents the exclusive authority by which the U.S. government may use, modify, distribute, or disclose the Software. The EULA and the license set forth therein, does not require or permit, among other things, that Keysight: (1) Furnish technical information related to commercial computer software or commercial computer software documentation that is not customarily provided to the public; or (2) Relinquish to, or otherwise provide, the government rights in excess of these rights customarily provided to the public to use, modify, reproduce, release, perform, display, or disclose commercial computer software or commercial computer software documentation. No additional government requirements beyond those set forth in the EULA shall

apply, except to the extent that those terms, rights, or licenses are explicitly required from all providers of commercial computer software pursuant to the FAR and the DFARS and are set forth specifically in writing elsewhere in the EULA. Keysight shall be under no obligation to update, revise or otherwise modify the Software. With respect to any technical data as defined by FAR 2.101, pursuant to FAR 12.211 and 27.404.2 and DFARS 227.7102, the U.S. government acquires no greater than Limited Rights as defined in FAR 27.401 or DFARS 227.7103-5 (c), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

1. General Information	
Introduction	5
Description	5
Initial Inspection	6
Repackaging for Shipment.	7
Keysight Packing Materials	7
Non-Keysight Packing Materials	7
Specifications	8
DC Resistance	8
Operating Conditions	9
Dimensions	9
Weight	9
Supplemental Performance Characteristics	10
Temperature Coefficient and Stability	10
Storage Conditions	10
2. Calibration and Repair	
Introduction	11
Calibration	11
Calibration Interval.	14
Repair	14

Contents

1 General Information

Introduction

This chapter provides the 42030A Four-Terminal Pair Standard Resistor Set description, specifications, and related general information.

Description

The 42030A Four-Terminal Pair Standard Resistor Set consists of nine precision resistor standards. The set is constructed using high stability resistors and are virtually unaffected by changes in ambient temperature, ensuring measurement repeatability and reliability. **Table 1-1** lists the 42030A contents.

Table 1-1 42030A Contents

Model No. or Keysight Part No.	Description	Quantity
42031A	1 m Ω Standard Resistor	1
42032A	10 m Ω Standard Resistor	1
42033A	100 m Ω Standard Resistor	1
42034A	1 Ω Standard Resistor	1
42035A	10 Ω Standard Resistor	1
42036A	100 Ω Standard Resistor	1
42037A	1 k Ω Standard Resistor	1
42038A	10 k Ω Standard Resistor	1
42039A	100 k Ω Standard Resistor	1
42030-60100	Carrying Case	1
Option ABA	Operation and Service Manual	1

Table 1-1

42030A Contents (Continued)

Model No. or Keysight Part No.	Description	Quantity
Not Assigned	Calibration Report	1

The 42030A is designed for use in calibrating the Keysight LCR meters which have four-terminal pair measurement terminals, and can be directly connected to the measurement terminals. The LCR meter documentation will give detailed information on using the 42030A.

Initial Inspection

Inspect the shipping container for damage. If the shipping container or cushioning material is damaged, it should be kept until the shipping contents have been checked for completeness and the instrument has been checked mechanically and electrically. The shipping contents should consist of the items shown in **Table 1-1**. If the shipment is incomplete, or if there is mechanical damage or other defects, notify your nearest Keysight Technologies office. If the shipping container is damaged, or the cushioning material shows signs of stress, notify the carrier as well as the Keysight Technologies office. Keep the shipping materials for the carrier's inspection. Your Keysight Technologies office will arrange for repair or replacement, without waiting for the claim settlement.

Repackaging for Shipment

Keysight Packing Materials

Containers and materials identical to those used in factory packaging are available from Keysight Technologies. If the instrument is being returned to Keysight Technologies for servicing, attach a tag indicating the type of service required, return address, model number, and full serial number. Also, mark the container *FRAGILE* to ensure careful handling. In any correspondence, refer to the unit by model number and full serial number.

Non-Keysight Packing Materials

The following general instructions should be used for re-packing with commercially available packing materials:

1. If shipping to Keysight Technologies office or service center, attach a tag indicating the type of service required, return address, model number, and full serial number.
2. Use a strong shipping container. A double-walled carton made of 350 pound test material is adequate.
3. Use enough shock absorbing material (3- to 4-inch layer) around all sides of the case to provide a firm cushion and to prevent movement inside the container.
4. Seal the shipping container securely.
5. Mark the shipping container *FRAGILE* to ensure careful handling.
6. In any correspondence, refer to the unit by model number and full serial number.

Specifications

This section gives the complete specifications for the 42030A Four-Terminal Pair Standard Resistor Set. When the 42030A is shipped from the factory, it meets the specifications listed in this section.

DC Resistance

Table 1-2 lists the nominal DC resistance for each resistance module of the 42030A.

Table 1-2

DC Resistance

Model Number	DC Resistance ¹
42031A	1 m Ω \pm 0.2%
42032A	10 m Ω \pm 0.2%
42033A	100 m Ω \pm 0.2%
42034A	1 Ω \pm 0.2%
42035A	10 Ω \pm 0.1%
42036A	100 Ω \pm 0.1%
42037A	1 k Ω \pm 0.1%
42038A	10 k Ω \pm 0.1%
42039A	100 k Ω \pm 0.1%

1. Specified under ambient environmental conditions of: Temperature 23 °C \pm 5 °C,
Relative Humidity \leq 70%

Operating Conditions

The 42030A must be operated under the ambient environmental conditions listed in **Table 1-3**.

Table 1-3 Operating Conditions

Temperature	23 °C ± 5 °C
Relative Humidity	≤70%

Dimensions

Table 1-4 shows the 42030A dimensions.

Table 1-4 Dimensions

Resistance Module	94 mm (W) × 31 mm (H) × 67 mm (D)
Carrying Case	400 mm (W) × 120 mm (H) × 290 mm (D)

Weight

Table 1-5 shows the 42030A weight.

Table 1-5 Weight

Resistance Module	170 g (ea.)
42030A Total	3,700 g

Supplemental Performance Characteristics

This section gives the supplemental performance characteristics for the 42030A Four-Terminal Pair Standard Resistor Set. These supplemental performance characteristics are not guaranteed, they are the 42030A typical characteristics that may be expected.

Temperature Coefficient and Stability

Table 1-6 shows the 42030A DC resistance temperature coefficient and stability.

Table 1-6 DC Resistance Temperature Coefficient and Stability

Model Number	Nominal Value	Temperature Coefficient ¹	Stability ²
42031A	1 mΩ	±15 ppm/°C	±100 ppm/year
42032A	10 mΩ	±15 ppm/°C	±100 ppm/year
42033A	100 mΩ	±15 ppm/°C	±100 ppm/year
42034A	1 Ω	±15 ppm/°C	±100 ppm/year
42035A	10 Ω	±10 ppm/°C	±100 ppm/year
42036A	100 Ω	±10 ppm/°C	±100 ppm/year
42037A	1 kΩ	±10 ppm/°C	±100 ppm/year
42038A	10 kΩ	±10 ppm/°C	±100 ppm/year
42039A	100 kΩ	±10 ppm/°C	±100 ppm/year

1. Under temperature condition: 23 °C ± 5 °C

2. Under environmental conditions of: Temperature 23 °C ± 5 °C,
 Relative Humidity ≤70%

Storage Conditions

The 42030A must be stored or shipped under the ambient environmental conditions listed in **Table 1-7**.

Table 1-7 Storage Conditions

Temperature	-40 °C to +70 °C
Relative Humidity	≤95% at 40 °C

2 Calibration and Repair

Introduction

This chapter provides the calibration and repair information for the 42030A Four-Terminal Pair Standard Resistor Set.

Calibration

The 42030A is calibrated in the parameters given in [Table 2-1](#) with the calibration uncertainties given in [Table 2-2](#) through [Table 2-7](#) and equipped with a calibration report, when shipped from the factory.

Table 2-1 42030A Calibration Parameters

Model Number	Nominal Value	DC Resistance	Series Resistance and Reactance	Parallel Resistance and Susceptance
42031A	1 m Ω	✓		
42032A	10 m Ω	✓		
42033A	100 m Ω	✓		
42034A	1 Ω	✓		
42035A	10 Ω	✓	✓	
42036A	100 Ω	✓	✓	
42037A	1 k Ω	✓		✓
42038A	10 k Ω	✓		✓
42039A	100 k Ω	✓		✓

Table 2-2

42030A DC Resistance Calibration Uncertainty

Model Number	Nominal Value	Calibration Uncertainty
42031A	1 mΩ	± 0.1%
42032A	10 mΩ	± 0.1%
42033A	100 mΩ	± 0.1%
42034A	1 Ω	± 0.1%
42035A	10 Ω	± 0.03%
42036A	100 Ω	± 0.02%
42037A	1 kΩ	± 0.02%
42038A	10 kΩ	± 0.02%
42039A	100 kΩ	± 0.02%

Table 2-3

42035A 10 Ω Resistor AC Calibration Uncertainty

Frequency	Calibration Uncertainty	
	Series Resistance	Reactance
1 MHz	± 0.12%	± 0.010 Ω
2 MHz	± 0.15%	± 0.010 Ω
3 MHz	± 0.20%	± 0.015 Ω
4 MHz	± 0.25%	± 0.025 Ω
5 MHz	± 0.4%	± 0.03 Ω
10 MHz	± 1.0%	± 0.08 Ω
13 MHz	± 1.5%	± 0.10 Ω

Table 2-4

42036A 100 Ω Resistor AC Calibration Uncertainty

Frequency	Calibration Uncertainty	
	Series Resistance	Reactance
1 MHz	± 0.12%	± 0.10 Ω
2 MHz	± 0.12%	± 0.10 Ω
3 MHz	± 0.15%	± 0.15 Ω
4 MHz	± 0.15%	± 0.15 Ω

Table 2-4 42036A 100 Ω Resistor AC Calibration Uncertainty (Continued)

Frequency	Calibration Uncertainty	
	Series Resistance	Reactance
5 MHz	$\pm 0.20\%$	$\pm 0.20 \Omega$
10 MHz	$\pm 0.5\%$	$\pm 0.5 \Omega$
13 MHz	$\pm 0.7\%$	$\pm 0.7 \Omega$

Table 2-5 42037A 1 k Ω Resistor AC Calibration Uncertainty

Frequency	Calibration Uncertainty	
	Parallel Resistance	Susceptance
100 kHz	$\pm 0.12\%$	$\pm 0.8 \mu\text{S}$
1 MHz	$\pm 0.08\%$	$\pm 1.0 \mu\text{S}$
2 MHz	$\pm 0.10\%$	$\pm 1.0 \mu\text{S}$
3 MHz	$\pm 0.10\%$	$\pm 1.5 \mu\text{S}$
4 MHz	$\pm 0.15\%$	$\pm 2.0 \mu\text{S}$
5 MHz	$\pm 0.15\%$	$\pm 3.0 \mu\text{S}$
10 MHz	$\pm 0.4\%$	$\pm 5 \mu\text{S}$
13 MHz	$\pm 0.6\%$	$\pm 8 \mu\text{S}$

Table 2-6 42038A 10 k Ω Resistor AC Calibration Uncertainty

Frequency	Calibration Uncertainty	
	Parallel Resistance	Susceptance
100 kHz	$\pm 0.08\%$	$\pm 0.10 \mu\text{S}$
1 MHz	$\pm 0.08\%$	$\pm 0.20 \mu\text{S}$

Table 2-7 42039A 100 k Ω Resistor AC Calibration Uncertainty

Frequency	Calibration Uncertainty	
	Parallel Resistance	Susceptance
100 kHz	$\pm 0.08\%$	$\pm 0.02 \mu\text{S}$
1 MHz	$\pm 0.12\%$	$\pm 0.15 \mu\text{S}$

Calibration and Repair Repair

The 42030A must be calibrated at a facility which satisfies the calibration uncertainties given in [Table 2-2](#) through [Table 2-7](#). Keysight Technologies' calibration laboratories satisfy all of these calibration uncertainties. For complete information on calibration, contact your nearest Keysight Technologies sales and service office.

Calibration Interval

The 42030A's recommended calibration interval is one year.

Repair

The 42030A can be repaired only at Keysight Technologies facilities. For complete information on repair, contact your nearest Keysight Technologies Sales and Service Office.

This information is subject to change without notice.
© Keysight Technologies 1992 - 2019
Edition 3, August 28, 2019



42030-90001
www.keysight.com